



INSPECTION TEST SIEVE VERIFICATION





Based upon the data below, Gilson Company, Inc. **VERIFIES** that this sieve meets the requirements set forth by ASTM in Standard Specification for Woven Wire Test Sieve Cloth and Test Sieves Designation: E11-09, **for Inspection Test Sieves.**

The data for this sieve was derived using a View Voyager 1212 optical imaging device, following internal work instruction # **WI-750-006**. The View -Voyager 1212 machine, serial # 361 was last **calibrated on 9/22/2011** with calibration **due on 10/01/2012**. This optical calibration was performed using calibration standard, serial number: 91284 and is traceable to N.I.S.T by file No. 821/272649-06

New certification
 Re-certification
 Sieve Visually Inspected
 Sieve cleaned

Wire Diameter

Opening Size



Bonnie Tucker

Laboratory Technician

12-5-11

Date

Joh Miller

Reviewed By

12-5-11

Date



P 740.548.7298 / 800.444.1508
 F 740.548.5314 / 800.255.5314
 W www.globalgilson.com