



MATCHED SIEVE CERTIFICATION

Matching Material:	BRONZE POWDER BP-77-22	Date:	5/29/2013
U.S. Number:	100	Customer:	GLISON
Nominal Opening:	150UM	Customer Order #:	XXXXXX
Test Number:	0	Tyler Order Number:	XXXX
Sample Size:	<input type="text" value="100Grams"/>		

Match Lab Results

Master Sieve Serial Number:

Matched Sieve Serial Number:

Weight Retained:

Weight Retained:

Percent Retained:

Percent Retained:

The above "MATCHED" sieve has been performance tested to insure that it will perform within + or - 2.5% of the weight in grams retained value of the "MASTER" sieve.

Allowable Tolerance of Matched Sieve

to

W.S. Tyler certifies that the above sieve meets the requirement of ASTM E-11

Matched By:



Q.C. Technician

P 740.548.7298 / 800.444.1508

F 740.548.5314 / 800.255.5314

W www.globalgilson.com